

Air and surfaces purifier

User Manual

^{*} For your safety and optimal functioning of the device, please read this user manual carefully.

^{*} The appearance and specifications of the product may be changed without prior notice due to product improvements.



Air and surface purifier nuvohla room/lift/med

Product	Nuvohla
Model	Room / Lift / Med
Consumption	3.6Wh
Sound	Less than 40 dB
Protection area	30/50 m²
Usage location	Indoor use only
Weight	0,5kg
Size	204 x 147 x 65 mm (H x W x D)

Introduction

This device's technology has been developed following the theory of the "natural purifier" present in the atmosphere, known as hydroxyl radical OH. This OH technology provides a unique air and surfaces purification solution, continuously inhibiting indoor bacteria, viruses, fungi, odours and other volatile organic compounds. The performance of the **nuvohla** purifier promotes a high air quality level, protecting the whole family.

Before turning on the device, carefully read this user manual to get all the information you know for precises and safety.

Index

- 1. Product features
- 2. Technology
- 3. Components
- 4. Warnings and precautions
- 5. Installation and assembly
- 6. Operating modes
- 7. Replacement and spare parts
- 8. Checks before contacting the after-sales service SAT
- 9. Warranty
- 10. Product identification

1. Product features

• Powerful purification and deodorisation

Hydroxyl radicals produce a chemical reaction which provides a unique active air and Surface purification solution. It deactivates bacteria viruses, fungi, odours and VOCs indoors.

· Constant high air quality level

Indoor pollution is produced by human activity and continuously added from outside. People need clean air. **nuvohla** purifier has been developed to operate 24 hours a day and guarantee the high air quality required by living beings.

Proven technology

The effectiveness of this technology regarding the inhibition of viruses and bacteria has been tested in independent laboratories. In addition, case studies have been carried out in hospitals, educational establishments and vet clinics, among others.

• Easy to use and economical to maintain

The maintenance required for **nuvohla** is the replacement of the cartridge when it runs out. The lifetime of the cartridge varies depending on the heat/humidity conditions of the environment. The electrical power consumption of the device is low. **nuvohla** does not require filters or UV lamps.

Contamination level detection

Depending on the indoor air quality, a LED-light will show different colours: blue, green, orange and red.

Error detection system

In case the cartridge is incorrectly inserted or is not inside the device, a flashing and two fix red-light signal will be activated.

• Cartridge replacement alarm signal

One week before the cartridge runs out, the replacement indicator will be activated and can be replaced. When 3 months have elapsed since cartridge installation, the red-light indicator will be activated and the device will automatically stop.


Additional features

- Wall and surfaces mounting pack available
- Wide coverage: the area of effect of the **nuvohla** purifying device is 30-50 m² depending on the contamination and environmental conditions


2. Technology

Hydroxyl radicals are generated from sunlight and are called "nature's detergent". These natural substances are easily extinguished in the environment with other contaminants after inactivation.


How does hydroxyl radical work?



Step 1: Hydroxyl radicals surround the protective coating of bacteria and viruses.




Step 2: Hydroxyl radicals react with hydrogen from bacteria or virus protective coating.




Step 3: Converts pollutants into H₂O.

3. Components

nuvohla pack consists of:




nuvohla Room / Med




nuvohla Lift

4. Installation

On surface installation. Mounting of the foot of the unit (Room / Med models only).




Wall installation. Mounting the support.




Left image corresponds to nuvohla Room / Med and right image to nuvohla Lift.

After mounting the bracket (stand or wall support), the power supply must be connected to the device and to the main socket.




In case of nuvohla Room / Med, finally, to start up the equipment, press the power button.




5. Warnings and precautions

WARNING: Improper use of this device can result in serious injury (including death) such as electric shock, fire, or damage to the device. Follow the instructions in this manual carefully.






Unplug the device and call the service centre when the cable or plug is damaged.



Check the socket is free of water and/or dust before plug it.




Do not bend the power supply cable or place any objects on top of it.



Keep the device away from fire, candles, heaters, chemicals, abrasives, fumes, smoke, sprays, butane gas or any other compound that may cause a deflagration or explosion.




Do not use the device with extreme humidity conditions.



Stop the device and disconnect it from the power supply when an unusual noise or odour is detected. Switch off and disconnect the device in case of catastrophic events such as flood, fire, etc.

CAUTION: Improper use of this device may result in property damage, malfunction or poor performance. Carefully follow the instructions in this manual.



supply before handling or moving it.

away from any light source, avoid

direct sunlight.



When replacing empty cartridges, they may contain a minimal amount of purifying solution. Please DO NOT ingest the contents and avoid ingestion by other living beings.



Please, discard empty cartridge properly into the recycling container indicated in accordance with the regulations of each country.

6. Operating modes

6.1 Room / Med

Air and surfaces purifier **nuvohla room / med** has three air quality sensor settings and two operating modes: ON/OFF mode and Night mode.

On the upper half front of the device, you will find a horizontal RGB light signal indicating the air quality status.

ON mode

When the device is connected to the power supply and the power button is pressed, the "warm up time" of the air quality sensor is activated for 3 minutes. The ON/OFF LED indicator lights blue and the front RGB LED changes its intensity in blue (or remains off depending on configuration 1-2-3). If during the warm up time the night mode is activated, the front RGB LED turns off and the fan speed is reduced.

After this time, the RGB light signal will stop the latency (except in night mode which remains off) and will work as explained in each configuration.

Night mode:

Night mode light signal turns GREEN. All other lights are turned off. Operates at lower speed and reduces the sound level.


Cartridge change:

The cartridge change light signal is RED. It lights up one week before it runs out. The RGB light signal operates normally depending on each configuration.


Alarm warning:

In case the cartridge runs out or is not inside the device, the RED error light signal will remain. In addition, the RGB led will be flashing in red. The green or blue LED and the device stops.

On the right side there are the 4 signal lights indicating:



CONFIGURATION 1 (factory default)



^{*} According to substances present in the air: perfumes, colognes, deodorants, alcohol, hydroalcoholic gel, tobacco smoke, etc.

CONFIGURATION 2

Disables the sensor for volatile organic compounds (air pollutants) and the front RGB light signal always remains blue.


CONFIGURATION 3

Disables the sensor for volatile organic compounds (air pollutants) and the front RGB light signal remains off.

HOW TO CHANGE THE CONFIGURATION

To find out in which configuration the device is operating, press and hold the Night mode icon until the side indicator lights up. To change the configuration, press and hold the night mode for at least 5 seconds, and while holding down the night mode, press the ON/OFF button as many times as necessary to choose the desired mode:

- Configuration 1 red light signal (error alarm).
- Configuration 2 red light signal (error alarm) + red light signal (cartridge change).
- Configuration 3 green light signal (night mode) + red light signal (cartridge change) + red light signal (error alarm).



Power failure

After a power failure, when power is restored, the device resumes the configuration and mode in which it was previously. For example: if the device is in configuration 2 and night mode and there is a power failure, when the power is restored, it will return to configuration 2 and night mode.

In addition, it always does the 3-minute warm up time, even if it is in night mode or in configuration 3 when it is not visible.

6.2 Lift

The device **nuvohla lift** has only one operating and configuration mode.

Once connected to the power supply, it switches on and starts working with the RGB led in blue and at high constant speed.

Cartridge change:

A week before the cartridge runs out, the RGB led changes colour to orange. Then, starts the alarm warning.

Alarm warning:

If the device doesn't detect a cartridge, there is no cartridge inside or the cartridge is empty, the unit shows an error warning with the RGB LED flashing red and cartridge change and alarm warning LEDs turn on.


7. Replacements and spare parts

Nuvohla purifier works thanks to the cartridge inserted in the unit. It must be inserted into the space indicated by lifting the non-adhesive end of the label and leaving this side facing outwards for easy removal. After three months of operation with the same cartridge, the device emits a light signal indicating that the cartridge is empty and must be replaced¹.

How to remove the cartridge?

The **nuvohla room/lift/med** device has two accesses to the cartridge:


- Rear access: a small access door that opens outwards by pressing the upper tab.



The cartridge is removed by grasping the label with your fingers and pulling it out of the device.

¹ Even if the tank contains some liquid, replace the cartridge anyway, otherwise your purification device will not perform its functions.

- Front access: at the base of the device there are two Phillips screws that need to be unscrewed so that the front cover of the device can be removed by pulling it off from the bottom. An access door like the one on the back of the device is then visible. The process of changing the cartridge is the same.




To insert the cartridge and close the device, follow the process in reverse.

How to clean the pre-filter?

The **nuvohla** purification device has a foam pre-filter which protects the fan and the other inside components from dust or insects. This pre-filter must be cleaned every time the cartridge is changed, or once a month if you are in a dusty environment.

To access the pre-filter, open the side door of the device by pressing its lateral flap and opening it to the left. The pre-filter can then be removed. The pre-filter is cleaned by shaking it slightly and then rinsing it with clean water. Let it dry completely before inserting it back into the device.



8. Failures and solutions

Before contacting the customer service, please check all the points listed below in detail. If you do not find the solution to your problem on this list, we will be able to find a solution to your problem.

Most common problems and solutions		
SYMPTOM	PROBLEM/FAILURE	SOLUTION
No light is illuminated	Is the power supply connected?	Correctly connect the transformer included in the box with the device to the power supply and to the device (p. 8).
	In case the device is the room/med model, is it switched on?	Press ON button
Many LEDs light but the fan does	Non original or expired cartridge	Insert an original cartridge supplied by an authorised dealer of nuvohla
not rotate	Incorrectly inserted cartridge	Open one of the cartridge access doors and make sure the cartridge is correctly positioned.
Air quality indicator is	Configuration 2 selected	Change configuration (operation section p. 14)
always blue	The air quality parameters are correct	The device is working correctly, but the air quality is very good and there is no change in quality
	It is a Lift device	The Lift model always displays the RGB LED in blue unless there is a warning (p. 15).
Air quality indicator does not light up	Configuration 3 selected	Change configuration (operation section p. 14)
The air quality indicator is always red	The device is too close to a source of pollution	Change the location of the device.

9. Warranty

Our warranty is valid for two years from the date of delivery (invoice). The guarantee is limited to the repair or replacement from our warehouse of products which may be faulty due to production defects. Installation costs shall be borne by the purchaser. Products which in our opinion have been improperly used, incorrectly handled, improperly installed, connected to voltages other than the rated voltage, modified or repaired by unauthorised personnel or which have been damaged during transport are excluded from all warranty claims.

10. Product identification



All **nuvohla** devices have their unique identification on its back, by means of a label that indicates:

- Unique serial number for each device, essential to facilitate spare parts or technical information.
- Product name
- Name and address of the manufacturer
- Power consumption
- Power supply

In case you detect errors or ambiguities on this manual, we will be pleased to receive your feedback, it helps us to improve our documentation even more. Airtècnics reserves the right to change any of the specifications in this manual.